

Human tPA Total Antigen Assay

Strip well format. Reagents for up to 96 tests.

For Research Use Only.

INTENDED USE

This human tissue plasminogen activator (tPA) total antigen assay is intended for the quantitative determination of human tPA total antigen in human biological fluids.

BACKGROUND

Tissue plasminogen activator is a serine protease that catalyzes the activation of plasminogen to plasmin [1]. Clinical studies have indicated that high tPA levels may increase the risk for thrombosis [2], whereas decreased levels may cause neuronal plasticity and degeneration [3].

ASSAY PRINCIPLE

tPA present in plasma or other biological fluids will react with the capture antibody coated on the microtiter plate. Free and complexed single and two-chain tPA will bind to the plate. After appropriate washing steps, polyclonal anti-human tPA primary antibody is added. Excess primary antibody is washed away and bound polyclonal antibody is then reacted with the horseradish peroxidase conjugated secondary antibody. Following an additional washing step, TMB substrate is used for color development at 450nm. The amount of color development is directly proportional to the original concentration of total tPA antigen in the sample.

REAGENTS PROVIDED

- ◆ **Immunoassay plate:**
1-96 well immulon plate coated, blocked, and dried with capture antibody
- ◆ **10X Wash Buffer:**
1 bottle of 50ml wash; bring to 1X using DI water
- ◆ **Human tPA antigen standard:**
1 vial of lyophilized standard
- ◆ **Anti-human tPA primary antibody:**
1 vial of lyophilized polyclonal antibody
- ◆ **Anti-rabbit horseradish peroxidase conjugated secondary antibody:**
1 vial of concentrated HRP labeled antibody
- ◆ **TMB substrate solution:**
1 bottle of 10ml solution

STORAGE AND STABILITY

All kit components must be stored at 4°C. Store unopened plate and any unused microtiter strips in the pouch with desiccant. Reconstituted standards and primary may be stored at -70°C for later use. **DO NOT** freeze/thaw the standards and primary antibody more than once. All other unused kit components must be stored at 4°C. Kit should be used no later than the expiration date.

REAGENTS AND EQUIPMENT REQUIRED

- 1-channel pipettes covering 0-10µl and 200-1000µl
- 12-channel pipette covering 30-300µl
- Paper towels or kimwipes
- 50ml tubes
- 1N H₂SO₄
- DI water
- Magnetic stirrer and stir-bars
- Plastic containers with lids

- TBS buffer
- 3% Blocking buffer
- Microtiter plate spectrophotometer operable at 450nm
- Microtiter plate shaker with uniform horizontally circular movement up to 300rpm.

WARNINGS

Warning - The tPA standards are of human origin. Each donor unit has been tested and found negative for the presence of HBsAg, anti-HIV 1+2, anti-HBc, and anti-HCV.

Since no tests are currently available to assure that no infectious agents are present, the standard must be treated as is recommended at the Biosafety Level 2 as potentially infectious human serum or blood specimen in the Centres for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories", 1984.

Warning – Avoid skin and eye contact when using TMB One substrate solution since it may be irritating to eyes, skin, and respiratory system. Wear safety goggles and gloves.

PRECAUTIONS

- **DO NOT** mix any reagents or components of this kit with any reagents or components of any other kit. This kit is designed to work properly as provided.
- **DO NOT** pipette reagents by mouth.
- Always pour substrate out of the bottle into a clean test tube. **DO NOT** pipette out of the bottle as you could contaminate the substrate.
- Keep plate covered except when adding reagents, washing, or reading.
- **DO NOT** smoke, drink, or eat in areas where specimens or reagents are being handled.

PREPARATION OF REAGENTS

- TBS buffer:** 0.1M Tris-HCl 0.15M NaCl, pH 7.4
- Blocking buffer:** 3% BSA in TBS buffer

SPECIMEN COLLECTION

Samples of human plasma, serum, urine, cell culture media, or tissue extracts may be applied directly to the plate.

The assay measures total human tPA in the 0.1-10 ng/ml range. Samples giving tPA levels above 10 ng/ml should be diluted in a similar biological fluid devoid of tPA. Blocking buffer may also be used.

ASSAY PROCEDURE

Perform assay at room temperature. Vigorously shake plate (300rpm) at each step of the assay.

Preparation of Standard:

Reconstitute standard as directed on vial to give a 1,000 ng/ml standard solution.

Dilution table for preparation of human tPA standard:

tPA concentration (ng/ml)	Dilutions
10	990µl (BSA) + 10µl (from vial)
5	500µl (BSA) + 500µl (10ng/ml)
2	600µl (BSA) + 400µl (5ng/ml)
1	500µl (BSA) + 500µl (2ng/ml)
0.5	500µl (BSA) + 500µl (1ng/ml)
0.25	500µl (BSA) + 500µl (0.5ng/ml)
0.1	600µl (BSA) + 400µl (0.25ng/ml)

NOTE: DILUTIONS FOR THE STANDARD CURVE MUST BE MADE AND APPLIED TO THE PLATE IMMEDIATELY.

Standard and Unknown Addition:

Remove microtiter plate from bag. Add 100µl of tPA standards (in duplicate) and unknowns to wells. Carefully record the position of standards and unknowns. Shake plate at 300rpm for 30 minutes. Wash wells three times with 300µl wash buffer. Remove excess wash by gently tapping plate on paper towel or kimwipe.

NOTE: If the unknown is thought to have high tPA levels, dilutions may be made in 3% BSA blocking buffer.

Primary Antibody Addition:

Reconstitute primary antibody by adding 10ml blocking buffer to vial. Agitate gently to completely dissolve contents. Add 100µl to all wells. Shake plate at 300rpm for 30 minutes. Wash wells three times with 300µl wash buffer. Remove excess wash by gently tapping plate on paper towel or kimwipe.

Secondary Antibody Addition:

Dilute 1µl conjugated secondary antibody in 10ml BSA blocking buffer and add 100µl to all wells. Shake plate at 300rpm for 30 minutes. Wash wells three times with 300µl wash buffer. Remove excess wash by gently tapping plate on paper towel or kimwipe.

Substrate Incubation:

Add 100µl of substrate solution to all wells and shake plate for 2-10 minutes. Quench the reaction with the addition of 50µl of 1N H₂SO₄ to all wells and read final absorbance values at 450nm.

NOTE: Time for substrate development is dependent on needs of researcher.



Measurement:

Set the absorbance at 450nm in a microtiter plate spectrophotometer. Measure the absorbance in all wells at 450nm, A₄₅₀.

Assay Calibration:

Plot A₄₅₀ against the amount of tPA in the standards. Fit a straight line through the points using a linear fit procedure. The tPA activity in the unknowns can be determined from this curve.

A typical standard curve.
(EXAMPLE ONLY, DO NOT USE)



EXPECTED VALUES

The basal level of tPA in healthy males and females, age 25-34 years, were found to be 5.5ng/ml and 4.0ng/ml respectively. tPA antigen increases with age; males and females age 55-64 years the median levels are 8.6ng/ml and 7.6ng/ml respectively [9,10].

Abnormalities in tPA levels have been reported in the following conditions:

- ◆ Neuronal plasticity and degeneration: Decreased levels of tPA have been implicated in the process of neuronal plasticity and degeneration [1,3].

- ◆ Arthritis: Decreased tPA levels may exacerbate arthritis [4].
- ◆ Deep venous thrombosis: Increased tPA levels may contribute to deep venous thrombosis [2].
- ◆ Coronary heart disease: Increased tPA levels may contribute to severe coronary heart disease [2].
- ◆ Pregnancy: Increased tPA levels are observed during pregnancy [7].
- ◆ Myocardial infarction: Increased tPA levels are observed in men who suffered MI compared to the matched controls [11].
- ◆ Stroke: Increased tPA levels have been associated with future stroke [12].

PERFORMANCE CHARACTERISTICS

The assay measures total tPA in the 0.1-10 ng/ml range. Samples giving tPA levels above 10 ng/ml should be diluted in 3% BSA blocking buffer.

DISCLAIMER

This information is believed to be correct but does not claim to be all-inclusive and shall be used only as a guide. The supplier of this kit shall not be held liable for any damage resulting from handling or from contact with the above product.

REFERENCES

1. Sallés Fernando J, *et al.*: Localization and Regulation of the Tissue Plasminogen Activator – Plasmin System in the Hippocampus. *The Journal of Neuroscience.*, March;**22(6)**: 2125-2134, 2002.
2. Wiman Björn, *et al.*: Inactivation of Tissue Plasminogen Activator in Plasma. *The Journal of Biological Chemistry*, March;**259(6)**: 3644-47, 1984.
3. Hastings Gregg A., *et al.*: Neuroserpin, a Brain-associated Inhibitor of Tissue Plasminogen Activator

Is Localized Primarily in Neurons. *The Journal of Biological Chemistry*, Dec;**272(52)**: 33062-33067, 1997.

4. Yang YH, *et al.*: Tissue-type Plasminogen Activator Deficiency Exacerbates Arthritis. *J Immunol*, July15;**167(2)**: 1047-52, 2001.

5. Chmielewska J, *et al.*: Determination of Tissue Plasminogen Activator and its "fast" Inhibitor in Plasma. *Clin Chem*, Mar**32(3)**: 482-5, 1986.

6. Ranby M, *et al.*: Blood Collection in Strong Acidic Citrate Anticoagulant used in a Study of Dietary Influence on Basal tPA

Activity. *Thromb Haemost*, Nov24;**62(3)**: 917-22, 1989.

7. Kruithof EK, *et al.*: Fibrinolysis in Pregnancy: a study of plasminogen activator inhibitors. *Blood*, Feb;**69(2)**: 460-466, 1987.

8. Eliasson, *et al.*: Influence of Gender, Age and Sampling Time on Plasma Fibrinolytic Variables and Fibrinogen. *Fibrinolysis* **7**: 316-323, 1993.

9. Ranby, M *et al.*: Age dependence of tissue plasminogen activator concentrations plasma, as studied by an improved enzyme-linked immunosorbent assay. *Clin Chem Dec*;**32(12)**: 2160-5, 1986.

10. Sundell, IB *et al.*: Fibrinolytic variables are related to age, sex, blood pressure, and body build measurements: a cross-sectional study in Norsjö, Sweden. *J Clin Epidemiol* **42(8)**: 719-23, 1989.

11. Ridker, PM *et al.*: Endogenous tissue-type plasminogen activator and risk of myocardial infarction. *Lancet* May 8; **341(8854)**: 1165-8, 1993.

12. Ridker, PM *et al.*: Prospective study of endogenous tissue plasminogen activator and risk of stroke. *Lancet* **343**: 940-943, 1994.